





Test Report

(Electronic version)

No:21R002085

Resistance to penetration	by biologically	y contamina [.]	ted aero	sols	M	
(ISO/DIS 22611:2003 Tempe	erature: 24, 5°C.	Rh:56.0%	Sample p	retreatment:time	of	roo

6.62

(ISO/DIS 22611:2003, Temperature:24.5°C, Rh:56.0%, Sample pretreatment:time of root temperature 48h)

Specimen 1 Thickness (mm) 0.18 Average mass per unit area(g/m^2) 110 Penetration ratio 6.60 Specimen 2 Thickness (mm) Average mass per unit area(g/m²) 110 6.53 Penetration ratio Specimen 3 0.18 Thickness (mm) Average mass per unit area(g/m²) 110 Penetration ratio 6.71 Specimen 4 0.18 Thickness (mm) Average mass per unit area(g/m^2) 110 Penetration ratio 6.63

Average penetration ratio

Requirement Penetration ratio >5; Class 3 was pass (EN 14126:2003/AC:2004)

---End of Report-

Page (GF)

电话:020-61994598/61994599 电话:020-37721161/66348638







Test Report

(Electronic version)

No: 21R002085



总部:广州市番禺区珠江路1号 花都实验室:广州市花都区狮岭镇旗岭河滨西路1号

电话:020-61994598/61994599 电话:020-37721161/66348638







Test Report

(Electronic version)

Verification Code: VGBH-3839-34 Verification Website: www.gttc.net.cn

No:21R002085 - X- Issue Date: 2021-10-08

Applicant: Wujiang Tutaike Textiles & Finishing Co.,LTD

Address: NO.1599 South 3rd Ring Road Shengze Wujiang Jlangsu

Information confirmed by applicant:

TTK-C01 Protective reusable apron coat

Quantity: 10 pieces

Manufacture's name: Wujiang Tutaike Textiles & Finishing Co.,LTD

Standard Adopted:

EN 14126:2003/AC:2004 < Protective clothing - Performance requirements and tests methods for protective clothing against infective agents>

Date Received/Date Test Started: 2021-09-25

Conclusion:

Resistance to penetration by biologically contaminated aerosols

M

Note: "M"-Meet the standard's requirement "F"-Fail to meet the standard's requirement "---"-No comment

Remark:

All the tested items are tested under the standard condition (except for indication).

Copies of the report are valid only re-stamped.

The experiment was carried out at No.1, Zhujiang Road, Panyu District, Guangzhou, Guangdong, P.R.China.

Approved By:

Liam Liu

LiQin Liu Senior Engineer

